

Calibrating Unit New Standard Solution Report

Sol. #204
6/10/15


Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01266 Calib. Date: 06/10/2015 Calib. No.: 00031
Certification File No.: 01267 Cert. Date: 06/10/2015 Cert. No.: 00022
Linearity File No.: 01268 Lin. Date: 06/10/2015 Lin. No.: 00022
Solution File No.: 01269 Soln. Date: 06/10/2015 Soln. No.: 00204
Sequential File No.: 01269 File Date: 06/10/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040
Control Solution %: 0.100% Expires: 02/14/2016
Solution Control Lot: 14B127 Bottle No.: 0822

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:11D		
Control 1 EC	0.101%	11:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:13D		
Control 2 EC	0.101%	11:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14D		
Control 3 EC	0.100%	11:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-444 

Changed By:

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: SGT JS/A # 6027

Badge No.: 6027

Date: 06/10/2015