

**ORIGINAL**

# Calibrating Unit New Standard Solution Report

SOL #175  
6/16/15

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01604  
Certification File No.: 01605  
Linearity File No.: 01606  
Solution File No.: 01655  
Sequential File No.: 01655

Serial No.: ARAJ-0075

Calib. Date: 04/07/2015      Calib. No.: 00018  
Cert. Date: 04/07/2015      Cert. No.: 00013  
Lin. Date: 04/07/2015      Lin. No.: 00013  
Soln. Date: 06/16/2015      Soln. No.: 00175  
File Date: 06/16/2015

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 02/11/2017  
Bottle No.: 0922

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:26D		
Control 1 EC	0.098%	21:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:27D		
Control 2 EC	0.098%	21:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	21:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:28D		
Control 3 EC	0.097%	21:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	21:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: Kenneth J Thomas # 314

Badge No.: 314

Date: 06/16/2015