

ORIGINAL

Sol. # 209
7/4/15

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01266
Certification File No.: 01267
Linearity File No.: 01268
Solution File No.: 01278
Sequential File No.: 01278

Serial No.: ARWF-0394
Calib. No.: 00031
Cert. No.: 00022
Lin. No.: 00022
Soln. No.: 00205
File Date: 07/04/2015

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34
Serial No.: DDAN-0040
Expires: 02/11/2017
Bottle No.: 0919

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:56D		
Control 1 EC	0.100%	16:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:57D		
Control 2 EC	0.100%	16:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:59D		
Control 3 EC	0.100%	16:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: _____

Kenneth / Thom #314

Badge No.: 314

Date: 07/04/2015