

**ORIGINAL**

# Calibrating Unit New Standard Solution Report

Sol. # 179  
7/20/15

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 01604 Calib. Date: 04/07/2015 Calib. No.: 00018  
Certification File No.: 01605 Cert. Date: 04/07/2015 Cert. No.: 00013  
Linearity File No.: 01606 Lin. Date: 04/07/2015 Lin. No.: 00013  
Solution File No.: 01678 Soln. Date: 07/20/2015 Soln. No.: 00179  
Sequential File No.: 01678 File Date: 07/20/2015

---

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041  
Control Solution %: 0.100% Expires: 02/11/2017  
Solution Control Lot: 15B137 Bottle No.: 0911

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:37D		
Control 1 EC	0.099%	16:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:38D		
Control 2 EC	0.099%	16:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:40D		
Control 3 EC	0.098%	16:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MAIRE

First Name: GUY

MI: E

Signature: \_\_\_\_\_

*JEM #282*

Badge No.: 1282

Date: 07/20/2015