

Calibrating Unit New Standard Solution Report

ORIGINAL

Sol. #181

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01604
Certification File No.: 01605
Linearity File No.: 01606
Solution File No.: 01693
Sequential File No.: 01693

Serial No.: ARAJ-0075

8/8/15

Calib. Date: 04/07/2015
Cert. Date: 04/07/2015
Lin. Date: 04/07/2015
Soln. Date: 08/08/2015
File Date: 08/08/2015

Calib. No.: 00018
Cert. No.: 00013
Lin. No.: 00013
Soln. No.: 00181

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 02/11/2017
Bottle No.: 0914

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:09D		
Control 1 EC	0.098%	00:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:10D		
Control 2 EC	0.098%	00:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:11D		
Control 3 EC	0.098%	00:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SYSOL

First Name: TIM

MI: M

Signature: _____



Badge No.: 289

Date: 08/08/2015