

ORIGINAL

**Calibrating Unit
New Standard Solution Report**

*Solo # 206
8/8/15*

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01266
Certification File No.: 01267
Linearity File No.: 01268
Solution File No.: 01293
Sequential File No.: 01293

Serial No.: ARWF-0394
Calib. Date: 06/10/2015
Cert. Date: 06/10/2015
Lin. Date: 06/10/2015
Soln. Date: 08/07/2015
File Date: 08/07/2015
Calib. No.: 00031
Cert. No.: 00022
Lin. No.: 00022
Soln. No.: 00206

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34
Serial No.: DDAN-0040
Expires: 02/11/2017
Bottle No.: 0924

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:55D		
Control 1 EC	0.101%	23:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:57D		
Control 2 EC	0.100%	23:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:58D		
Control 3 EC	0.100%	23:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: _____

[Handwritten Signature] #282

Badge No.: 1282

Date: 08/08/2015