

Calibrating Unit New Standard Solution Report

ORIGINAL

SOLC # 185

4/4/15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D. -
Calibration File No.: 01604
Certification File No.: 01605
Linearity File No.: 01606
Solution File No.: 01725
Sequential File No.: 01725

Serial No.: ARAJ-0075

Calib. Date: 04/07/2015
Cert. Date: 04/07/2015
Lin. Date: 04/07/2015
Soln. Date: 09/04/2015
File Date: 09/04/2015

Calib. No.: 00018
Cert. No.: 00013
Lin. No.: 00013
Soln. No.: 00185

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 02/11/2017
Bottle No.: 0917

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:53D		
Control 1 EC	0.099%	10:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:55D		
Control 2 EC	0.099%	10:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56D		
Control 3 EC	0.098%	10:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: _____



Badge No.: 1282

Date: 09/04/2015