

D-100-001

Calibrating Unit New Standard Solution Report

Sol # 186

9-12-15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01604
Certification File No.: 01605
Linearity File No.: 01606
Solution File No.: 01730
Sequential File No.: 01730

Serial No.: ARAJ-0075

Calib. Date: 04/07/2015 Calib. No.: 00018
Cert. Date: 04/07/2015 Cert. No.: 00013
Lin. Date: 04/07/2015 Lin. No.: 00013
Soln. Date: 09/12/2015 Soln. No.: 00186
File Date: 09/12/2015

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 02/11/2017
Bottle No.: 0912

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:47D		
Control 1 EC	0.099%	20:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:48D		
Control 2 EC	0.099%	20:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:49D		
Control 3 EC	0.099%	20:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: _____

Kenneth Thomas #314

Badge No.: 314

Date: 09/12/2015