

ORIGINAL

# Calibrating Unit New Standard Solution Report

Sol # 187

9-17-15

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 01604  
 Certification File No.: 01605  
 Linearity File No.: 01606  
 Solution File No.: 01732  
 Sequential File No.: 01732

Calib. Date: 04/07/2015  
 Cert. Date: 04/07/2015  
 Lin. Date: 04/07/2015  
 Soln. Date: 09/17/2015  
 File Date: 09/17/2015

Serial No.: ARAJ-0075  
 Calib. No.: 00018  
 Cert. No.: 00013  
 Lin. No.: 00013  
 Soln. No.: 00187

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041  
 Expires: 02/11/2017  
 Bottle No.: 0902

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:51D		
Control 1 EC	0.100%	18:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:52D		
Control 2 EC	0.099%	18:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:54D		
Control 3 EC	0.098%	18:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: \_\_\_\_\_

*Kenneth Thom #314*

Badge No.: 314

Date: 09/17/2015