

# Calibrating Unit New Standard Solution Report

ORIGINAL  
Sol# 188  
9/28/15

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01604  
Certification File No.: 01605  
Linearity File No.: 01606  
Solution File No.: 01741  
Sequential File No.: 01741

Serial No.: ARAJ-0075

Calib. Date: 04/07/2015  
Cert. Date: 04/07/2015  
Lin. Date: 04/07/2015  
Soln. Date: 09/28/2015  
File Date: 09/28/2015

Calib. No.: 00018  
Cert. No.: 00013  
Lin. No.: 00013  
Soln. No.: 00188

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 02/11/2017  
Bottle No.: 0908

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:00D		
Control 1 EC	0.099%	10:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 2 EC	0.098%	10:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		
Control 3 EC	0.098%	10:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: \_\_\_\_\_



Badge No.: 1282

Date: 09/28/2015