

ORIGINAL

Calibrating Unit New Standard Solution Report

Sol-190
10-23-15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01746
Certification File No.: 01747
Linearity File No.: 01748
Solution File No.: 01757
Sequential File No.: 01757

Serial No.: ARAJ-0075

Calib. Date: 10/02/2015
Cert. Date: 10/02/2015
Lin. Date: 10/02/2015
Soln. Date: 10/23/2015
File Date: 10/23/2015

Calib. No.: 00019
Cert. No.: 00014
Lin. No.: 00014
Soln. No.: 00190

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 02/11/2017
Bottle No.: 0904

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:56D		
Control 1 EC	0.099%	14:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57D		
Control 2 EC	0.098%	14:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58D		
Control 3 EC	0.098%	14:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: _____

Kenneth Thomas #314

Badge No.: 314

Date: 10/23/2015