

# Calibrating Unit New Standard Solution Report

ORIGINAL  
Sol # 193  
11/19/15

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01746  
Certification File No.: 01747  
Linearity File No.: 01748  
Solution File No.: 01773  
Sequential File No.: 01773

Serial No.: ARAJ-0075

Calib. Date: 10/02/2015  
Cert. Date: 10/02/2015  
Lin. Date: 10/02/2015  
Soln. Date: 11/19/2015  
File Date: 11/19/2015

Calib. No.: 00019  
Cert. No.: 00014  
Lin. No.: 00014  
Soln. No.: 00193

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 04/13/2017  
Bottle No.: 0214

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:22S		
Control 1 EC	0.099%	23:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:23S		
Control 2 EC	0.098%	23:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:24S		
Control 3 EC	0.098%	23:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SYSOL

First Name: TIM

MI: M

Signature: \_\_\_\_\_



Badge No.: 289

Date: 11/19/2015