

Calibrating Unit New Standard Solution Report

RECEIVED
 Sol. #
 194
 11/28/15

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01746
 Certification File No.: 01747
 Linearity File No.: 01748
 Solution File No.: 01780
 Sequential File No.: 01780

Calib. Date: 10/02/2015
 Cert. Date: 10/02/2015
 Lin. Date: 10/02/2015
 Soln. Date: 11/28/2015
 File Date: 11/28/2015

Serial No.: ARAJ-0075
 Calib. No.: 00019
 Cert. No.: 00014
 Lin. No.: 00014
 Soln. No.: 00194

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDAN-0041
 Expires: 04/13/2017
 Bottle No.: 0208

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 04:36S | | |
| Control 1 EC | 0.099% | 04:37S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 04:37S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:38S | | |
| Control 2 EC | 0.097% | 04:38S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 04:38S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:39S | | |
| Control 3 EC | 0.096% | 04:39S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 04:39S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:40S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SYSOL

First Name: TIM

MI: M

Signature: _____

Sysol #289

Badge No.: 289

Date: 11/28/2015