

Calibrating Unit New Standard Solution Report

801-195
12-10-15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01746
Certification File No.: 01747
Linearity File No.: 01748
Solution File No.: 01784
Sequential File No.: 01784

Serial No.: ARAJ-0075
Calib. No.: 00019
Cert. No.: 00014
Lin. No.: 00014
Soln. No.: 00195
File Date: 12/10/2015

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15D138

Model No.: CU-34
Serial No.: DDAN-0041
Expires: 04/13/2017
Bottle No.: 0203

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:09S		
Control 1 EC	0.099%	01:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:11S		
Control 2 EC	0.098%	01:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:12S		
Control 3 EC	0.097%	01:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	01:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SYSOL

First Name: TIM

MI: M

Signature: _____



Badge No.: 289

Date: 12/10/2015