

Calibrating Unit

New Standard Solution Report

ORIGINAL

501.212

Equipment
Location: Alcotest 7110 MKIII-C
TOMS RIVER P.D.
Calibration File No.: 01324
Certification File No.: 01325
Linearity File No.: 01326
Solution File No.: 01348
Sequential File No.: 01348

Serial No.: ARWF-0394

12/23/15

Calib. Date: 10/02/2015
Cert. Date: 10/02/2015
Lin. Date: 10/02/2015
Soln. Date: 12/23/2015
File Date: 12/23/2015

Calib. No.: 00032
Cert. No.: 00023
Lin. No.: 00023
Soln. No.: 00212

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDAN-0040
Expires: 04/13/2017
Bottle No.: 0201

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:00S		
Control 1 EC	0.097%	12:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	12:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01S		
Control 2 EC	0.097%	12:02S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:03S		
Control 3 EC	0.097%	12:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: _____



Badge No.: 1282

Date: 12/23/2015