

ORIGINAL

Calibrating Unit New Standard Solution Report

sol. #
201
1/15/16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01746
Certification File No.: 01747
Linearity File No.: 01748
Solution File No.: 01806
Sequential File No.: 01806

Serial No.: ARAJ-0075
Calib. Date: 10/02/2015
Cert. Date: 10/02/2015
Lin. Date: 10/02/2015
Soln. Date: 01/15/2016
File Date: 01/15/2016

Calib. No.: 00019
Cert. No.: 00014
Lin. No.: 00014
Soln. No.: 00201

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 04/13/2017
Bottle No.: 0215

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:56S		
Control 1 EC	0.099%	14:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57S		
Control 2 EC	0.095%	14:58S	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.092%	14:58S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:59S		
Control 3 EC	0.095%	14:59S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.094%	14:59S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:00S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: _____

Kenneth Thomas #314

Badge No.: 314

Date: 01/15/2016