

Calibrating Unit

New Standard Solution Report

SOL# 214
2-4-16


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01324	Calib. Date: 10/02/2015
Certification File No.:	01325	Cert. No.: 00032
Linearity File No.:	01326	Lin. Date: 10/02/2015
Solution File No.:	01370	Lin. No.: 00023
Sequential File No.:	01370	Soln. Date: 02/04/2016
		Soln. No.: 00214
		File Date: 02/04/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	15D138	Expires: 04/13/2017
		Bottle No.: 0219

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:25S		
Control 1 EC	0.099%	23:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:26S		
Control 2 EC	0.099%	23:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:28S		
Control 3 EC	0.100%	23:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS	First Name: KENNETH	MI: J
Signature: <u> #314</u>	Badge No.: 314	Date: 02/04/2016