

ORIGINAL

Calibrating Unit New Standard Solution Report

8/10/218
3/10/16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01324
Certification File No.: 01325
Linearity File No.: 01326
Solution File No.: 01395
Sequential File No.: 01395

Serial No.: ARWF-0394
Calib. No.: 00032
Cert. No.: 00023
Lin. No.: 00023
Soln. No.: 00218

Calib. Date: 10/02/2015
Cert. Date: 10/02/2015
Lin. Date: 10/02/2015
Soln. Date: 03/10/2016
File Date: 03/10/2016

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDWK S3-0387
Expires: 04/13/2017
Bottle No.: 0217

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:16S		
Control 1 EC	0.092%	23:16S	33.8°C	CONTROL TEST FAILED
Control 1 IR	0.093%	23:16S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	23:26S		
Control 2 EC	0.095%	23:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	23:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:27S		
Control 3 EC	0.095%	23:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.095%	23:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:29S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: _____

Kenneth Thomas #314

Badge No.: 314

Date: 03/10/2016