

Calibrating Unit New Standard Solution Report



ORIGINAL

Sol. # 206

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01818
Certification File No.: 01819
Linearity File No.: 01820
Solution File No.: 01822
Sequential File No.: 01822

Serial No.: ARAJ-0075 9/20/16

Calib. Date: 03/18/2016
Cert. Date: 03/18/2016
Lin. Date: 03/18/2016
Soln. Date: 04/20/2016
File Date: 04/20/2016

Calib. No.: 00021
Cert. No.: 00015
Lin. No.: 00015
Soln. No.: 00206

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 04/13/2017
Bottle No.: 0210

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:13D		
Control 1 EC	0.103%	16:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:14D		
Control 2 EC	0.101%	16:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15D		
Control 3 EC	0.101%	16:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature: *Adam Koepen* # 350

Badge No.: 350

Date: 04/20/2016