

Calibrating Unit New Standard Solution Report

 ORIGINAL
Sol # 207

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01818
Certification File No.: 01819
Linearity File No.: 01820
Solution File No.: 01834
Sequential File No.: 01834

Serial No.: ARAJ-0075 5-1-16

Calib. Date: 03/18/2016 Calib. No.: 00021
Cert. Date: 03/18/2016 Cert. No.: 00015
Lin. Date: 03/18/2016 Lin. No.: 00015
Soln. Date: 05/01/2016 Soln. No.: 00207
File Date: 05/01/2016

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15D138

Model No.: CU-34
Serial No.: DDWK S3-0386
Expires: 04/13/2017
Bottle No.: 0222

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:14D		
Control 1 EC	0.100%	02:15D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:16D		
Control 2 EC	0.100%	02:16D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:17D		
Control 3 EC	0.100%	02:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:18D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: 

Badge No.: 306

Date: 05/01/2016