

Calibrating Unit New Standard Solution Report

Sol. # 222
5/2/16

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01407 Calib. Date: 03/18/2016 Calib. No.: 00033
Certification File No.: 01408 Cert. Date: 03/18/2016 Cert. No.: 00024
Linearity File No.: 01409 Lin. Date: 03/18/2016 Lin. No.: 00024
Solution File No.: 01438 Soln. Date: 05/02/2016 Soln. No.: 00222
Sequential File No.: 01438 File Date: 05/02/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387
Control Solution %: 0.100% Expires: 03/22/2018
Solution Control Lot: 201603L Bottle No.: 0412

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:51D		
Control 1 EC	-.---%	12:52D	22.0°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	12:52D	22.0°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature: 

350

Badge No.: 350

Date: 05/02/2016