

# Calibrating Unit New Standard Solution Report

Sol. # 223

5/2/16

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01407 Calib. Date: 03/18/2016 Calib. No.: 00033  
Certification File No.: 01408 Cert. Date: 03/18/2016 Cert. No.: 00024  
Linearity File No.: 01409 Lin. Date: 03/18/2016 Lin. No.: 00024  
Solution File No.: 01439 Soln. Date: 05/02/2016 Soln. No.: 00223  
Sequential File No.: 01439 File Date: 05/02/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387  
Control Solution %: 0.100% Expires: 02/22/2018  
Solution Control Lot: 201603L Bottle No.: 0416

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:55D		
Control 1 EC	0.101%	14:56D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56D		
Control 2 EC	0.100%	14:57D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58D		
Control 3 EC	0.101%	14:58D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #350

Badge No.: 350

Date: 05/02/2016