

# Calibrating Unit

## New Standard Solution Report

So/ #208

5/23/16

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01818  
Certification File No.: 01819  
Linearity File No.: 01820  
Solution File No.: 01845  
Sequential File No.: 01845

Calib. Date: 03/18/2016  
Cert. Date: 03/18/2016  
Lin. Date: 03/18/2016  
Soln. Date: 05/23/2016  
File Date: 05/23/2016

Serial No.: ARAJ-0075

Calib. No.: 00021  
Cert. No.: 00015  
Lin. No.: 00015  
Soln. No.: 00208

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 201603L

Model No.: CU-34

Serial No.: DDWK S3-0386  
Expires: 03/22/2018  
Bottle No.: 0808

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 09:33D |                |                     |
| Control 1 EC      | 0.099% | 09:34D | 33.9°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 09:34D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:35D |                |                     |
| Control 2 EC      | 0.099% | 09:35D | 33.9°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 09:35D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:36D |                |                     |
| Control 3 EC      | 0.099% | 09:36D | 33.9°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.099% | 09:36D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:37D |                |                     |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature: 

Badge No.: 350

Date: 05/23/2016