

Calibrating Unit New Standard Solution Report

Sol. # 210
6/20/16

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01818
 Certification File No.: 01819
 Linearity File No.: 01820
 Solution File No.: 01876
 Sequential File No.: 01876

Serial No.: ARAJ-0075

Calib. Date: 03/18/2016 Calib. No.: 00021
 Cert. Date: 03/18/2016 Cert. No.: 00015
 Lin. Date: 03/18/2016 Lin. No.: 00015
 Soln. Date: 06/20/2016 Soln. No.: 00210
 File Date: 06/20/2016

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 201603

Model No.: CU-34 Serial No.: DDWK S3-0386
 Expires: 03/22/2018
 Bottle No.: 0809

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:02D		
Control 1 EC	0.100%	16:02D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:03D		
Control 2 EC	0.099%	16:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:04D		
Control 3 EC	0.099%	16:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

[Handwritten Signature] *[Initials]* *[Sol]*

Badge No.: 306

Date: 06/20/2016