

Calibrating Unit New Standard Solution Report

ORIGINAL

Sol #
225
7/13/16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01407
Certification File No.: 01408
Linearity File No.: 01409
Solution File No.: 01466
Sequential File No.: 01466

Calib. Date: 03/18/2016
Cert. Date: 03/18/2016
Lin. Date: 03/18/2016
Soln. Date: 07/13/2016
File Date: 07/13/2016

Serial No.: ARWF-0394
Calib. No.: 00033
Cert. No.: 00024
Lin. No.: 00024
Soln. No.: 00225

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 201603L

Model No.: CU-34

Serial No.: DDWK S3-0387
Expires: 03/22/2018
Bottle No.: 0414

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:13D		
Control 1 EC	0.100%	15:14D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14D		
Control 2 EC	0.100%	15:15D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:16D		
Control 3 EC	0.099%	15:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 07/13/2016