

Item 15 of 19

Alcotest 7110 Calibration Certificate

ARAJ-0075

Part I - Control Tests

Equipment

Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01934
 Certification File No.: 01935
 Linearity File No.: 01820
 Solution File No.: 01933
 Sequential File No.: 01935

Serial No.: ARAJ-0075

Calib. Date: 10/14/2016
 Cert. Date: 10/14/2016
 Lin. Date: 03/18/2016
 Soln. Date: 09/30/2016
 File Date: 10/14/2016

Calib. No.: 00023
 Cert. No.: 00016
 Lin. No.: 00015
 Soln. No.: 00219

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 16220

Model No.: CU-34

Serial No.: DDAN-0041
 Expires: 09/07/2018
 Bottle No.: 0456

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:59D		
Control 1 EC	0.098%	11:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00D		
Control 2 EC	0.098%	12:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01D		
Control 3 EC	0.098%	12:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02D		

All tests within acceptable tolerance.

Coordinator

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature:

Tpr. I K W. #6704

Badge No.: 6704

Date: 10/14/2016

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.