

# Calibrating Unit New Standard Solution Report

Sol. #  
223

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 01934 Calib. Date: 10/14/2016 Calib. No.: 00023  
Certification File No.: 01935 Cert. Date: 10/14/2016 Cert. No.: 00016  
Linearity File No.: 01936 Lin. Date: 10/14/2016 Lin. No.: 00016  
Solution File No.: 01966 Soln. Date: 11/21/2016 Soln. No.: 00223  
Sequential File No.: 01966 File Date: 11/21/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041  
Control Solution %: 0.100% Expires: 06/07/2018  
Solution Control Lot: 16140 Bottle No.: 1481

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:56S		
Control 1 EC	0.099%	09:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58S		
Control 2 EC	0.098%	09:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59S		
Control 3 EC	0.099%	10:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:

 #350

Badge No.: 350

Date: 11/21/2016