

# Calibrating Unit New Standard Solution Report

JAN - 3 2017

*SOL # 228  
12/18/16*

**Equipment** Alcotest 7110 MKIII-C  
**Location:** TOMS RIVER P.D.  
**Calibration File No.:** 01985  
**Certification File No.:** 01986  
**Linearity File No.:** 01987  
**Solution File No.:** 01995  
**Sequential File No.:** 01995

**Calib. Date:** 12/14/2016  
**Cert. Date:** 12/14/2016  
**Lin. Date:** 12/14/2016  
**Soln. Date:** 12/18/2016  
**File Date:** 12/18/2016

**Serial No.:** ARAJ-0075  
**Calib. No.:** 00024  
**Cert. No.:** 00017  
**Lin. No.:** 00017  
**Soln. No.:** 00228

**Calibrating Unit:** WET  
**Control Solution %:** 0.100%  
**Solution Control Lot:** 16140

**Model No.:** CU-34

**Serial No.:** DDAN-0041  
**Expires:** 06/07/2018  
**Bottle No.:** 1212

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:44S		
Control 1 EC	0.097%	21:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:45S		
Control 2 EC	0.096%	21:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:47S		
Control 3 EC	0.097%	21:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	21:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NATER

First Name: MARK

MI: S

Signature: \_\_\_\_\_

*Mark J. Nater #336*

Badge No.: 336

Date: 12/18/2016