

Calibrating Unit
New Standard Solution Report

ORIGINAL

#229
Sol
12/30/16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01985
Certification File No.: 01986
Linearity File No.: 01987
Solution File No.: 02000
Sequential File No.: 02000

Calib. Date: 12/14/2016
Cert. Date: 12/14/2016
Lin. Date: 12/14/2016
Soln. Date: 12/30/2016
File Date: 12/30/2016

Serial No.: ARAJ-0075
Calib. No.: 00024
Cert. No.: 00017
Lin. No.: 00017
Soln. No.: 00229

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 06/07/2018
Bottle No.: 1396

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:25S		
Control 1 EC	0.099%	11:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27S		
Control 2 EC	0.098%	11:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28S		
Control 3 EC	0.098%	11:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: _____

A. Elie #282

Badge No.: 1282

Date: 12/30/2016