

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01985
 Certification File No.: 01986
 Linearity File No.: 01987
 Solution File No.: 02004
 Sequential File No.: 02004

Serial No.: ARAJ-0075
 Calib. No.: 00024
 Cert. No.: 00017
 Lin. No.: 00017
 Soln. No.: 00230
 File Date: 01/06/2017

SOL # 230
1-6-17

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 16140

Model No.: CU-34
 Serial No.: DDWK S3-0386
 Expires: 06/07/2018
 Bottle No.: 1470

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:09S		
Control 1 EC	0.100%	13:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:11S		
Control 2 EC	0.099%	13:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:12S		
Control 3 EC	0.100%	13:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJPZ-025 (KA)

Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: TP. I K W. Q #6704

Badge No.: 6704

Date: 01/06/2017