



ORIGINAL

# Calibrating Unit New Standard Solution Report

SOL# 232  
1-21-17

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01482  
Certification File No.: 01483  
Linearity File No.: 01484  
Solution File No.: 01511  
Sequential File No.: 01511

Calib. Date: 11/04/2016  
Cert. Date: 11/04/2016  
Lin. Date: 11/04/2016  
Soln. Date: 01/21/2017  
File Date: 01/21/2017

Serial No.: ARWF-0394  
Calib. No.: 00036  
Cert. No.: 00026  
Lin. No.: 00026  
Soln. No.: 00232

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDAN-0040  
Expires: 06/07/2018  
Bottle No.: 1205

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:59S		
Control 1 EC	0.098%	03:00S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:01S		
Control 2 EC	0.097%	03:01S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	03:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:02S		
Control 3 EC	0.097%	03:03S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: \_\_\_\_\_

Badge No.: 306

Date: 01/21/2017