

Calibrating Unit New Standard Solution Report

Sol. # 231
ORIGINAL
1/29/17

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01985
Certification File No.: 01986
Linearity File No.: 01987
Solution File No.: 02011
Sequential File No.: 02011

Serial No.: ARAJ-0075
Calib. Date: 12/14/2016
Cert. Date: 12/14/2016
Lin. Date: 12/14/2016
Soln. Date: 01/29/2017
File Date: 01/29/2017

Calib. No.: 00024
Cert. No.: 00017
Lin. No.: 00017
Soln. No.: 00231

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 06/07/2018
Bottle No.: 1187

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:44S		
Control 1 EC	0.101%	03:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:45S		
Control 2 EC	0.098%	03:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:46S		
Control 3 EC	0.098%	03:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 01/29/2017