

Calibrating Unit New Standard Solution Report

Sol. #
ORIGINAL 233



1/29/17

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01482
Certification File No.: 01483
Linearity File No.: 01484
Solution File No.: 01520
Sequential File No.: 01520

Serial No.: ARWF-0394
Calib. Date: 11/04/2016
Cert. Date: 11/04/2016
Lin. Date: 11/04/2016
Soln. Date: 01/29/2017
File Date: 01/29/2017
Calib. No.: 00036
Cert. No.: 00026
Lin. No.: 00026
Soln. No.: 00233

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16390

Model No.: CU-34
Serial No.: DDAN-0040
Expires: 12/02/2018
Bottle No.: 1367

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:42S		
Control 1 EC	0.099%	08:42S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:43S		
Control 2 EC	0.098%	08:44S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:45S		
Control 3 EC	0.098%	08:45S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 01/29/2017