

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	01985	Calib. Date: 12/14/2016
Certification File No.:	01986	Calib. No.: 00024
Linearity File No.:	01987	Cert. Date: 12/14/2016
Solution File No.:	02023	Cert. No.: 00017
Sequential File No.:	02023	Lin. Date: 12/14/2016
		Lin. No.: 00017
		Soln. Date: 02/17/2017
		Soln. No.: 00232
		File Date: 02/17/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	16140	Expires: 06/07/2018
		Bottle No.: 1188

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:29S		
Control 1 EC	0.100%	17:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:31S		
Control 2 EC	0.099%	17:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:32S		
Control 3 EC	0.099%	17:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature: _____

Adam Koepen #350

Badge No.: 350

Date: 02/17/2017