



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01482	Calib. Date: 11/04/2016
Certification File No.:	01483	Calib. No.: 00036
Linearity File No.:	01484	Cert. Date: 11/04/2016
Solution File No.:	01531	Lin. No.: 00026
Sequential File No.:	01531	Lin. Date: 11/04/2016
		Soln. No.: 00234
		File Date: 02/18/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	16390	Expires: 12/02/2018
		Bottle No.: 1288

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:35S		
Control 1 EC	0.100%	02:35S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:36S		
Control 2 EC	0.098%	02:37S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:38S		
Control 3 EC	0.098%	02:38S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 02/18/2017