

Calibrating Unit New Standard Solution Report

SOL # 235

3/2/2017

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01985
Certification File No.: 01986
Linearity File No.: 01987
Solution File No.: 02039
Sequential File No.: 02039

Calib. Date: 12/14/2016
Cert. Date: 12/14/2016
Lin. Date: 12/14/2016
Soln. Date: 03/02/2017
File Date: 03/02/2017

Serial No.: ARAJ-0075
Calib. No.: 00024
Cert. No.: 00017
Lin. No.: 00017
Soln. No.: 00235

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16390

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 12/02/2018
Bottle No.: 1297

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 17:43S | | |
| Control 1 EC | 0.100% | 17:43S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 17:43S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:44S | | |
| Control 2 EC | 0.099% | 17:45S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 17:45S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:45S | | |
| Control 3 EC | 0.099% | 17:46S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 17:46S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:47S | | |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #350

Badge No.: 350

Date: 03/02/2017