

Calibrating Unit

New Standard Solution Report



ORIGINAL

SOL # 235

3/5/17

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01482
 Certification File No.: 01483
 Linearity File No.: 01484
 Solution File No.: 01538
 Sequential File No.: 01538

Calib. Date: 11/04/2016
 Cert. Date: 11/04/2016
 Lin. Date: 11/04/2016
 Soln. Date: 03/05/2017
 File Date: 03/05/2017

Serial No.: ARWF-0394
 Calib. No.: 00036
 Cert. No.: 00026
 Lin. No.: 00026
 Soln. No.: 00235

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 16390

Model No.: CU-34

Serial No.: DDAN-0040
 Expires: 12/02/2018
 Bottle No.: 1268

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:27S		
Control 1 EC	0.100%	05:28S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:29S		
Control 2 EC	0.099%	05:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:30S		
Control 3 EC	0.099%	05:31S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 03/05/2017