

# Calibrating Unit New Standard Solution Report

Sol # 236

3/18/17

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01482 Calib. Date: 11/04/2016 Calib. No.: 00036  
Certification File No.: 01483 Cert. Date: 11/04/2016 Cert. No.: 00026  
Linearity File No.: 01484 Lin. Date: 11/04/2016 Lin. No.: 00026  
Solution File No.: 01557 Soln. Date: 03/18/2017 Soln. No.: 00236  
Sequential File No.: 01557 File Date: 03/18/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040  
Control Solution %: 0.100% Expires: 12/02/2018  
Solution Control Lot: 16390 Bottle No.: 1302

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:06S		
Control 1 EC	0.100%	02:07S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:08S		
Control 2 EC	0.098%	02:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:09S		
Control 3 EC	0.099%	02:10S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DEROSA

First Name: MICHAEL

MI: A

Signature: \_\_\_\_\_

Badge No.: 386

Date: 03/18/2017