

Alcotest 7110 Calibration Record

Equipment

Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02042
Certification File No.: 01986
Linearity File No.: 01987
Solution File No.: 02041
Sequential File No.: 02042

Serial No.: ARAJ-0075

Calib. Date: 04/18/2017
Cert. Date: 12/14/2016
Lin. Date: 12/14/2016
Soln. Date: 03/31/2017
File Date: 04/18/2017

Calib. No.: 00026
Cert. No.: 00017
Lin. No.: 00017
Soln. No.: 00236

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16220

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 09/07/2018
Bottle No.: 0609

Coordinator

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: TPF. I K W. A #6704

Badge No.: 6704

Date: 04/18/2017

*Black Key Temperature Probe Serial.....#

ΔΔXKPZ-396 (KA)

*Digital NIST Temperature Measuring System Serial.....#

150 649 672 (KA)

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.