

Calibrating Unit

New Standard Solution Report

238

4-7-17

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D. Serial No.: ARWF-0394
Calibration File No.: 01482 Calib. Date: 11/04/2016 Calib. No.: 00036
Certification File No.: 01483 Cert. Date: 11/04/2016 Cert. No.: 00026
Linearity File No.: 01484 Lin. Date: 11/04/2016 Lin. No.: 00026
Solution File No.: 01572 Soln. Date: 04/07/2017 Soln. No.: 00238
Sequential File No.: 01572 File Date: 04/07/2017

Calibrating Unit: WET Model No.: 41 Serial No.: DDAN-0040
Control Solution %: 0.100% Expires: 12/02/2018
Solution Control Lot: 16390 Bottle No.: 1262

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:49D		
Control 1 EC	0.099%	02:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:51D		
Control 2 EC	0.098%	02:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:52D		
Control 3 EC	0.097%	02:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:54D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DEROSA

First Name: MICHAEL

MI: A

Signature: 

Badge No.: 386

Date: 04/07/2017