

238

4-24-17

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. Date:	04/18/2017
Calibration File No.:	02042	Calib. No.:	00026
Certification File No.:	02043	Cert. Date:	04/18/2017
Linearity File No.:	02044	Lin. Date:	04/18/2017
Solution File No.:	02050	Soln. Date:	04/24/2017
Sequential File No.:	02050	File Date:	04/24/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWK S3-0386
Control Solution %:	0.100%			Expires:	12/02/2018
Solution Control Lot:	16390			Bottle No.:	1210


Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:08D		
Control 1 EC	0.098%	05:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:09D		
Control 2 EC	0.098%	05:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:10D		
Control 3 EC	0.098%	05:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WORTH	First Name: ADAM	MI: C
------------------	------------------	-------

Signature: 	Badge No.: 378
	Date: 04/24/2017