

SOL # 241

5/28/17

ORIGINAL

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
 Location: TOMS RIVER P.D.
 Calibration File No.: 02054 Calib. Date: 05/12/2017 Calib. No.: 00028
 Certification File No.: 02055 Cert. Date: 05/12/2017 Cert. No.: 00019
 Linearity File No.: 02056 Lin. Date: 05/12/2017 Lin. No.: 00019
 Solution File No.: 02069 Soln. Date: 05/28/2017 Soln. No.: 00241
 Sequential File No.: 02069 File Date: 05/28/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
 Control Solution %: 0.100% Expires: 12/02/2018
 Solution Control Lot: 16390 Bottle No.: 1328


Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:58D		
Control 1 EC	0.101%	08:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00D		
Control 2 EC	0.100%	09:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01D		
Control 3 EC	0.100%	09:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL First Name: ROBERT MI: S

Signature: 

Badge No.: 306
Date: 05/28/2017