

SOL # 242
6/3/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01582	Calib. Date:	04/18/2017
Certification File No.:	01583	Cert. Date:	04/18/2017
Linearity File No.:	01584	Lin. Date:	04/18/2017
Solution File No.:	01606	Soln. Date:	06/03/2017
Sequential File No.:	01606	File Date:	06/03/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	16390	Expires:	12/02/2018
		Bottle No.:	1264

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:16D		
Control 1 EC	0.098%	01:17D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:17D		
Control 2 EC	0.098%	01:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:19D		
Control 3 EC	0.098%	01:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:

 # 350

Badge No.: 350

Date: 06/03/2017