

SOL # 243  
6-16-17

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 02054  
Certification File No.: 02055  
Linearity File No.: 02056  
Solution File No.: 02089  
Sequential File No.: 02089

Serial No.: ARAJ-0075

Calib. Date: 05/12/2017  
Cert. Date: 05/12/2017  
Lin. Date: 05/12/2017  
Soln. Date: 06/16/2017  
File Date: 06/16/2017

Calib. No.: 00028  
Cert. No.: 00019  
Lin. No.: 00019  
Soln. No.: 00243

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 16390

Model No.: CU-34

Serial No.: DDWK S3-0386  
Expires: 12/02/2018  
Bottle No.: 1292

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:12D		
Control 1 EC	0.099%	15:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:13D		
Control 2 EC	0.098%	15:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14D		
Control 3 EC	0.099%	15:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHWARTZ First Name: STEVE

MI: -

Signature: \_\_\_\_\_

Badge No.: 293

Date: 06/16/2017