



ORIGINAL

Calibrating Unit New Standard Solution Report

8/28/17

246

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01582
 Certification File No.: 01583
 Linearity File No.: 01584
 Solution File No.: 01624
 Sequential File No.: 01624

Calib. Date: 04/18/2017
 Cert. Date: 04/18/2017
 Lin. Date: 04/18/2017
 Soln. Date: 08/28/2017
 File Date: 08/28/2017

Serial No.: ARWF-0394
 Calib. No.: 00037
 Cert. No.: 00027
 Lin. No.: 00027
 Soln. No.: 00246

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 17110

Model No.: CU-34

Serial No.: DDWK S3-0387
 Expires: 03/20/2019
 Bottle No.: 0695

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:30D		
Control 1 EC	0.098%	21:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:31D		
Control 2 EC	0.097%	21:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:32D		
Control 3 EC	0.097%	21:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 08/28/2017