

SOL # 251

# Calibrating Unit New Standard Solution Report



**ORIGINAL**

9-11-17

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 02054  
 Certification File No.: 02055  
 Linearity File No.: 02056  
 Solution File No.: 02145  
 Sequential File No.: 02145

Serial No.: ARAJ-0075

Calib. Date: 05/12/2017    Calib. No.: 00028  
 Cert. Date: 05/12/2017    Cert. No.: 00019  
 Lin. Date: 05/12/2017    Lin. No.: 00019  
 Soln. Date: 09/11/2017    Soln. No.: 00251  
 File Date: 09/11/2017

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 17110

Model No.: CU-34    Serial No.: DDWK S3-0386  
 Expires: 03/20/2019  
 Bottle No.: 0522

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:12D		
Control 1 EC	0.100%	14:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:13D		
Control 2 EC	0.100%	14:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:14D		
Control 3 EC	0.100%	14:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WESTFALD  
 Signature: 

FF First Name: ROBERT  
 SOL

MI: S

Badge No.: 306  
 Date: 09/11/2017