

SOL # 254

10/20/17

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 02147  
Certification File No.: 02148  
Linearity File No.: 02149  
Solution File No.: 02170  
Sequential File No.: 02170

Serial No.: ARAJ-0075  
Calib. No.: 00029  
Cert. No.: 00020  
Lin. No.: 00020  
Soln. No.: 00254  
File Date: 10/20/2017

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 17110

Model No.: CU-34  
Serial No.: DDWK S3-0386  
Expires: 03/20/2019  
Bottle No.: 0013

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:39D		
Control 1 EC	0.099%	21:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:40D		
Control 2 EC	0.098%	21:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:42D		
Control 3 EC	0.099%	21:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:43D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #350

Badge No.: 350

Date: 10/20/2017