

SOLN # 249

11/4/17

Calibrating Unit New Standard Solution Report

| | | | |
|-------------------------|-----------------------|-------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARWF-0394 |
| Location: | TOMS RIVER P.D. | Calib. No.: | 00038 |
| Calibration File No.: | 01631 | Cert. Date: | 09/15/2017 |
| Certification File No.: | 01632 | Lin. Date: | 09/15/2017 |
| Linearity File No.: | 01633 | Soln. Date: | 11/04/2017 |
| Solution File No.: | 01639 | File Date: | 11/04/2017 |
| Sequential File No.: | 01639 | | |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDWK S3-0387 |
| Solution Control Lot: | 17110 | Expires: | 03/20/2019 |
| | | Bottle No.: | 0505 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 13:57S | | |
| Control 1 EC | 0.099% | 13:58S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 13:58S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:59S | | |
| Control 2 EC | 0.098% | 13:59S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 13:59S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:00S | | |
| Control 3 EC | 0.098% | 14:01S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 14:01S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:01S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: *Pascal Gambardella #365*

Badge No.: 365

Date: 11/04/2017