

SOL # 257

11-22-17

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 02147  
Certification File No.: 02148  
Linearity File No.: 02149  
Solution File No.: 02190  
Sequential File No.: 02190

Calib. Date: 09/15/2017  
Cert. Date: 09/15/2017  
Lin. Date: 09/15/2017  
Soln. Date: 11/22/2017  
File Date: 11/22/2017

Serial No.: ARAJ-0075  
Calib. No.: 00029  
Cert. No.: 00020  
Lin. No.: 00020  
Soln. No.: 00257

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 17110

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 03/20/2019  
Bottle No.: 0099

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:47S		
Control 1 EC	0.100%	08:47S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 2 EC	0.099%	08:49S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49S		
Control 3 EC	0.100%	08:50S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: ΔΔXK PZ-436 (KA)

**Changed By:**

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpt. I AK W. Q #6704

Badge No.: 6704

Date: 11/22/2017