

SOL # 258

# Calibrating Unit New Standard Solution Report

12/14/17


**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 02147 Calib. Date: 09/15/2017 Calib. No.: 00029  
Certification File No.: 02148 Cert. Date: 09/15/2017 Cert. No.: 00020  
Linearity File No.: 02149 Lin. Date: 09/15/2017 Lin. No.: 00020  
Solution File No.: 02216 Soln. Date: 12/14/2017 Soln. No.: 00258  
Sequential File No.: 02216 File Date: 12/14/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041  
Control Solution %: 0.100% Expires: 03/20/2019  
Solution Control Lot: 17110 Bottle No.: 0607

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:22S		
Control 1 EC	0.101%	01:22S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:23S		
Control 2 EC	0.101%	01:23S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:24S		
Control 3 EC	0.100%	01:25S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**  
Last Name: DEROSA First Name: MICHAEL MI: A  
Signature:  Badge No.: 386  
Date: 12/14/2017